



精准检测
PRECISE TESTING

CERTIFICATE OF CONFORMITY

No.: PTC20031706001S-LD01

Applicant : KEDY TECH (GANZHOU) ELECTRONICS CO., LTD
Address : 2nd floor, no.15 workshop, standard factory building base, chengbei industrial zone, Anyuan county, Ganzhou city, Jiangxi province
Manufacturer : KEDY TECH (GANZHOU) ELECTRONICS CO., LTD
Address : 2nd floor, no.15 workshop, standard factory building base, chengbei industrial zone, Anyuan county, Ganzhou city, Jiangxi province
Product : Infrared thermometer
Trade Mark : KODYEE
Model : CF-818
Power Rating : Battery: DC 3.0V

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

Low Voltage Directive - 2014/35/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 60601-1: 2006 + A11:2011 + A1:2013 + A12:2014	PTC20031706001S-LD01

This certificate is part of the full test report(s) and should be read in conjunction with it. This certificate is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Dongguan Precise Testing Service Co., Ltd, this certificate is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.



Jacky Ou
Manager
Date: March 24, 2020

Precise Testing & Certification (Guangdong) Co., Ltd.
Building 1, No.6, Tongxin Road, Dongcheng Street,
Dongguan, Guangdong, China
Tel: 86 769 38808222 Web: www.ptc-testing.com
Mail: inquiry@ptc-testing.com





精准检测
PRECISE TESTING

CERTIFICATE OF CONFORMITY

No.: PTC20031706001E-EM01

Applicant : KEDY TECH (GANZHOU) ELECTRONICS CO., LTD
Address : 2nd floor, no.15 workshop, standard factory building base,
chengbei industrial zone, Anyuan county, Ganzhou city, Jiangxi
province
Manufacturer : KEDY TECH (GANZHOU) ELECTRONICS CO., LTD
Address : 2nd floor, no.15 workshop, standard factory building base,
chengbei industrial zone, Anyuan county, Ganzhou city, Jiangxi
province
Trade Mark : N/A
Product : Infrared thermometer
Model : CF-818

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

EMC Directive - 2014/30/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 55011:2016+A1:2017 EN 60601-1-2:2017 EN 61000-3-2:2014 EN 61000-3-3:2013	PTC20031706001E-EM01

This certificate is part of the full test report(s) and should be read in conjunction with it. This certificate is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Precise Testing & Certification (Guangdong) Co., Ltd. this certificate is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.



Jacky Ou
Manager
Date: March 23, 2020

Precise Testing & Certification (Guangdong) Co., Ltd.
Building 1, No.6, Tongxin Road, Dongcheng Street,
Dongguan, Guangdong, China
Tel: 86 769 38808222 Web: www.ptc-testing.com
Mail: inquiry@ptc-testing.com



TEST REPORT EN 60825-1:2014 Safety of laser products – Part 1: Equipment classification and requirements and user's guide	
Report reference No	RDG200316015-SF
Compiled by (+ signature)	Danny Chen <i>Danny Chen</i>
Approved by (+ signature)	Safety Engineer: Jerry Liu <i>Jerry Liu</i>
Date of issue	2020-04-01
Testing laboratory	Bay Area Compliance Laboratories Corp. (Shenzhen)
Address	6/F., West Wing, Third Phase of Wanli Industrial Building, Shihua Road, Futian Free Trade Zone, Shenzhen, Guangdong, China
Testing location	As above
Applicant	KEDI TECHNOLOGY ELECTRONICS CO.,LTD
Address	KEDI AVENUE,JIULONG New Village , ANYU COUNTY INDUSTRIAL PARK,GANZHOU CITY JIANGXI,China
Standard	EN 60825-1:2014
Test sample(s) received	2020-03-28
Test in period	2020-03-30
Procedure deviation	N/A
Non-standard test method	N/A
Summary of testing.....	The sample tested complies with requirements of IEC 60825-1: 2014 (Edition 3.0)
This test report is specially limited to the above client company and product model only. It may not be duplicated without prior written consent of Bay Area Compliance Laboratories Corp. (Shenzhen).	
Type of test object	Infrared thermometer
Trademark	N/A
Model/type reference	CF-818
Manufacturer.....	The same as applicant.
Address	The same as applicant.
Rating	3V  (Building in AAA 1.5V*2 battery)

